

E.2 SALIENT CHARACTERISTICS

Offerors will be evaluated to the extent they can provide the items (or equal) listed in the price schedule in Section B. The Government will award to the lowest priced, responsible, technically acceptable offeror.

Items must meet the following salient characteristics:

Overall Size:

Able to fit on mobile cart shelf, not to exceed 16" x 20" x 6".

Features

Indicated for orthopedic and arthroscopic procedures of the knee, shoulder, ankle, elbow, wrist, and hip.

Must be a compatible connection hub for Stryker Formula Shaver Handpieces (both with and without buttons).

Must be compatible with Stryker Formula 180 Handpieces.

Must be compatible with Stryker Small-Joint Shaver Handpiece.

Must be a compatible connection hub for Stryker SERFAS RF cutting and coagulation probes.

Must include footswitch to allow switching between SERFAS and Shaver, and for operating both items.

Must be compatible with other Stryker Firewire devices.

Should have option of wireless footswitch.

120VAC power supply

Includes operators manuals and maintenance manuals

1 year warranty on parts and labor on equipment